

CE

DECLARATION OF CONFORMITY

According Directive 98/79/EC on in vitro diagnostic medical devices, Annex III.

Manufacturer: BEIJING KEWEI CLINICAL DIAGNOSTIC REAGENT INC.

Address: No.7, Yan Qi He, Xi Yi Rd., Huai Rou District, Beijing, China.

European Representative: Lotus NL B.V.

Contact person: Peter E-mail: peter@lotusnl.com

Address: Koningin Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.

In Vitro Diagnostic Directive:

- COVID-19 Antigen Rapid Test Kit (Oral Saliva).

Category: Others.

Conformity assessment route: Declaration of Conformity IVDD Annex III

Applicable Standards:

ISO13485:2016

ISO 14971:2019

ENISO18113-1:2011

ENISO18113-2:2011

ENISO18113-3:2011

EN 13641:2002

ISO 15223-1:2016

EN 13612:2002

ISO23640:2015

EN 62366-1:2015

We, the manufacturer, herewith declare with sole responsibility that our product/s mentioned above meet/s the provisions of the Directive 98/79/EC of the European Parliament and of the Council on In-Vitro Diagnostic Medical Devices.

We agree to develop, implement and maintain a documented post-production monitoring process.

Signed:

Place Beijing, China.

European Representative:

Seal/Stamp:

Lotus NL B.V.



Name of authorized signatory: *Wang Baojun*

Position held in the company: General Manager

Seal/Stamp:

BEIJING KEWEI CLINICAL

DIAGNOSTIC REAGENT INC.

19/02/2021